

# METHODS AND APPLICATIONS

## Sequence-specific determination of protein and peptide concentrations by absorbance at 205 nm

Nicholas J. Anthis and G. Marius Clore\*

Laboratory of Chemical Physics, National Institute of Diabetes and Digestive and Kidney Diseases, National Institutes of Health, Bethesda, Maryland, 20892-0520

Received 1 February 2013; Revised 11 March 2013; Accepted 14 March 2013

DOI: 10.1002/pro.2253

Published online [proteinscience.org](http://proteinscience.org)

**Abstract:** Quantitative studies in molecular and structural biology generally require accurate and precise determination of protein concentrations, preferably via a method that is both quick and straightforward to perform. The measurement of ultraviolet absorbance at 280 nm has proven especially useful, since the molar absorptivity (extinction coefficient) at 280 nm can be predicted directly from a protein sequence. This method, however, is only applicable to proteins that contain tryptophan or tyrosine residues. Absorbance at 205 nm, among other wavelengths, has been used as an alternative, although generally using absorptivity values that have to be uniquely calibrated for each protein, or otherwise only roughly estimated. Here, we propose and validate a method for predicting the molar absorptivity of a protein or peptide at 205 nm directly from its amino acid sequence, allowing one to accurately determine the concentrations of proteins that do not contain tyrosine or tryptophan residues. This method is simple to implement, requires no calibration, and should be suitable for a wide range of proteins and peptides.

**Keywords:** protein; absorbance; UV; concentration; molecular biology; absorptivity; extinction coefficient

### Introduction

Accurate determination of protein concentration is essential for quantitative biochemical, biophysical, molecular, and structural biology studies. A wide range of spectrophotometric methods are available for doing so, each with its own advantages, disadvantages, and

specific requirements.<sup>1</sup> The absorbance of ultraviolet (UV) radiation by intrinsic chromophores is one commonly used method; particularly useful is absorbance at 280 nm ( $A_{280}$ ), which offers high specificity, as it arises strictly from tryptophan and tyrosine residues (and to a small extent from disulfide bonds if present). Thus, the molar absorptivity (extinction coefficient) for a protein at 280 nm ( $\epsilon_{280}$ ) can be accurately estimated directly from its amino acid sequence.<sup>2–4</sup> Absorbance at wavelengths other than 280 nm is also used less commonly, generally either in a non-sequence-specific manner or by calibrating absorbance data on a protein-by-protein basis. Calibration can be time-consuming and technically difficult, especially if it requires directly

---

Grant sponsors: Intramural Program of the NIH, NIDDK, and the Intramural AIDS Targeted Antiviral Program of the Office of the Director of the NIH (to GMC).

\*Correspondence to: G. Marius Clore, Laboratory of Chemical Physics, National Institute of Diabetes and Digestive and Kidney Diseases, National Institutes of Health, Bethesda, MD 20892-0520. E-mail: [mariusco@mail.nih.gov](mailto:mariusco@mail.nih.gov)

weighing lyophilized protein or peptide (which is confounded by the presence of any residual salt or water remaining in the dried protein). Other methods (e.g., Bradford, Lowry assays) that do not rely on intrinsic chromophores also require calibration and have the added disadvantage of requiring additional reagents and time. Thus, the ability to predict the molar absorptivity (e.g., at 280 nm) is advantageous in terms of both accuracy and efficiency.

If a protein contains no tyrosine or tryptophan residues, however,  $A_{280}$  cannot be used to determine the concentration of that protein in solution. One alternative is  $A_{205}$ , which arises primarily from the peptide bond.<sup>5,6</sup> Although the maximum absorbance of a protein actually occurs closer to 190 nm,  $A_{205}$  has been favored in part due to the technical limitations of measuring at lower wavelengths.<sup>6,7</sup> Still, whereas most common buffers and solutions in biological research are essentially transparent at 280 nm, many solutes will exhibit some absorbance at 205 nm (although this effect is even more pronounced at lower wavelengths).<sup>1,5,8</sup> Thus, one must consider, and carefully control for, any  $A_{205}$  stemming from the buffer alone, as described below. On the other hand, the much higher sensitivity of  $A_{205}$  relative to  $A_{280}$  offers an additional advantage that helps counteract this. The ratio of  $A_{205}$  to  $A_{280}$  is  $\sim 30$  on average, although it varies widely from protein to protein (and is virtually infinite for a protein lacking tryptophan and tyrosine residues).

Absorbance at 205 nm arises primarily from the peptide backbone. Thus, one can roughly estimate the concentration of a protein solution (in terms of  $\text{mg}\cdot\text{mL}^{-1}$ ) without any knowledge of the protein sequence. A commonly used absorptivity value is  $\epsilon_{205} = 31 \text{ mL}\cdot\text{mg}^{-1}\cdot\text{cm}^{-1}$ .<sup>1,5,8</sup> However, side chain absorbance at 205 nm is still significant; in particular, the aromatic side chains (tryptophan, phenylalanine, tyrosine, and histidine) all have a greater  $A_{205}$  than a single peptide bond. Considering this, Scopes<sup>6</sup> proposed using the  $A_{280}/A_{205}$  ratio for a given protein to estimate a more accurate  $\epsilon_{205}$  (in effect taking into account tryptophan and tyrosine side chains). However, this does not include the contributions of phenylalanine, histidine, methionine, arginine, or cysteine/cystine, all of which absorb significantly at 205 nm. Thus, by taking into account the contributions of individual side chains to the overall  $A_{205}$ , one can accurately and specifically (i.e., in a sequence-specific manner) determine the concentration of virtually any protein in solution, as we demonstrate in this report.

## Results and Discussion

We measured the absorbance of six proteins and two peptides at both 205 nm and 280 nm to compare the values at these two wavelengths. Calmodulin (CaM) is a 148-residue calcium-binding protein,<sup>9</sup> which has

a predicted molar absorptivity at 280 nm of  $\epsilon_{280} = 2980 \text{ M}^{-1}\cdot\text{cm}^{-1}$  from two tyrosine residues and no tryptophan residues. Measurements were performed on both apo CaM and CaM saturated with four calcium ions ( $4\text{Ca}^{2+}$ -CaM). Calmodulin-dependent kinase 1 (CaMK1), is a serine/threonine kinase that is activated by calcium-loaded CaM.<sup>10</sup> One CaMK1 construct used in this study corresponds to the kinase domain (residues 1–296) and has a predicted  $\epsilon_{280}$  of  $42,860 \text{ M}^{-1}\cdot\text{cm}^{-1}$  from four tryptophans and 14 tyrosines. The other CaMK1 construct (residues 299–320) is a short peptide corresponding to the CaM-binding domain of CaMK1, with a predicted  $\epsilon_{280}$  of  $5500 \text{ M}^{-1}\cdot\text{cm}^{-1}$  from one tryptophan. The other peptide used in this study is the 26-residue M13 peptide derived from skeletal muscle myosin light-chain kinase (skMLCK),<sup>11</sup> which also has a predicted  $\epsilon_{280}$  of  $5500 \text{ M}^{-1}\cdot\text{cm}^{-1}$ , based on one tryptophan. The remaining four proteins measured were the B1 domain of Streptococcal protein G (GB1), the talin2 F3 domain, maltose-binding protein (MBP), and Enzyme I. These range in size from 66 to 575 residues in length, and have predicted  $\epsilon_{280}$  values ranging from  $9970 \text{ M}^{-1}\cdot\text{cm}^{-1}$  to  $66,350 \text{ M}^{-1}\cdot\text{cm}^{-1}$ , stemming from multiple tryptophan and tyrosine residues.

For greatest accuracy, a wide range of half-log dilutions was measured, though in everyday practice such extensive measurement would not be necessary or even warranted. All dilutions were performed in water. The stocks of Enzyme I and the two short peptides were already in water, but the stocks of the other proteins used in this study were in other buffers, some with significant absorbance at 205 nm. In these cases, dilutions were also made of the buffer alone, so that its absorbance could be subtracted from that of the protein solution. Thus, although the buffers used for CaMK1<sub>1–296</sub> and  $4\text{Ca}^{2+}$ -CaM, for example, had undiluted  $A_{205}$  values of 23.74 and 27.45, respectively, we were still able to accurately and precisely measure the  $A_{205}$  of CaMK1<sub>1–296</sub> (uncorrected = 678.2, corrected = 654.5) and CaM (uncorrected = 665.6, corrected = 638.2) using water dilutions in the range of 1:1000 to 1:10,000 (Table I). The buffer used for apo CaM also had a relatively high absorbance of  $A_{205} = 31.21$ , but the other buffers had much lower absorbances (0.59–4.37). Regardless, buffer absorbance corrections were applied in all of these cases. However, no such corrections were needed for skMLCK, CaMK1<sub>299–320</sub>, or Enzyme I, as these stock solutions were already in water. Although one could instead ensure that solutes free of absorbance at 205 nm are used, our results show that this is unnecessary. In practice (as done here) one would ideally measure  $A_{205}$  values much greater for the protein than for the buffer solution alone, but at the very least, the difference between the two measurements must be significantly larger than the experimental error.

**Table I.** Analysis of Data Collected for This Study

Protein	MW (g·mol <sup>-1</sup> ) <sup>a</sup>	$\epsilon_{280}$ (M <sup>-1</sup> ·cm <sup>-1</sup> ) <sup>b</sup>		$A_{280}$ <sup>c</sup>	$A_{205}$		Conc. ( $\mu$ M) <sup>d</sup>	$\epsilon_{bb}$ (M <sup>-1</sup> ·cm <sup>-1</sup> ) <sup>e</sup>
					Measured <sup>f</sup>	Corrected <sup>g</sup>		
CaMK1 <sub>299-320</sub>	2599	5500	Average <sup>h</sup>	9.72	170.97	170.97	1767.96	2,704
			Error <sup>i</sup>	0.97	5.36	5.36	176.49	482
			% Error <sup>j</sup>	9.98	3.13	3.13	9.98	17.8
skMLCK	2964	5500	Average <sup>h</sup>	7.90	162.85	162.85	1436.12	2,838
			Error <sup>i</sup>	0.63	8.63	8.63	114.70	435
			% Error <sup>j</sup>	7.99	5.30	5.30	7.99	15.3
GB1	7246	9970	Average <sup>h</sup>	5.72	152.40	151.80	573.53	3112
			Error <sup>i</sup>	0.25	8.07	8.07	25.22	281
			% Error <sup>j</sup>	4.40	5.30	5.32	4.40	9.0
talin2	11,558	16,960	Average <sup>h</sup>	18.35	433.25	432.66	1081.99	2870
			Error <sup>i</sup>	0.84	17.39	17.39	49.50	241
			% Error <sup>j</sup>	4.57	4.01	4.02	4.57	8.4
apo CaM	16,706	2890	Average <sup>h</sup>	2.65	488.01	457.81	889.23	2716
			Error <sup>i</sup>	0.09	39.26	39.26	30.87	324
			% Error <sup>j</sup>	3.47	8.05	8.58	3.47	11.9
4Ca <sup>2+</sup> -CaM	16,706	2890	Average <sup>h</sup>	3.86	665.60	638.15	1294.76	2567
			Error <sup>i</sup>	0.03	37.45	37.45	11.60	199
			% Error <sup>j</sup>	0.90	5.63	5.87	0.90	7.8
CaMK1 <sub>1-296</sub>	33,442	42,860	Average <sup>h</sup>	25.25	678.19	654.45	589.03	2631
			Error <sup>i</sup>	1.53	28.53	28.53	35.62	281
			% Error <sup>j</sup>	6.05	4.21	4.36	6.05	10.7
MBP	40,614	66,350	Average <sup>h</sup>	1.81	45.37	41.00	27.21	2914
			Error <sup>i</sup>	0.04	1.82	1.82	0.61	204
			% Error <sup>j</sup>	2.25	4.02	4.45	2.25	7.0
Enzyme I	63,467	24,410	Average <sup>h</sup>	28.84	2267.58	2267.58	1181.66	2684
			Error <sup>i</sup>	1.15	21.38	21.38	46.96	137
			% Error <sup>j</sup>	3.97	0.94	0.94	3.97	5.1

<sup>a</sup> Molecular weight calculated from amino acid sequence; confirmed by mass spectrometry.

<sup>b</sup> Molar absorptivity at 280 nm, calculated as described in Methods section.

<sup>c</sup> Absorbance at 280 nm of undiluted protein solution, extrapolated from measurements on various dilutions.

<sup>d</sup> Concentration calculated from absorbance at 280 nm.

<sup>e</sup> Molar absorptivity at 205 nm per backbone peptide bond, calculated individually for each protein as described in the text.

<sup>f</sup> Absorbance at 205 nm of undiluted protein solution, extrapolated from measurements on various dilutions.

<sup>g</sup> Absorbance at 205 nm of buffer alone was subtracted from measurements on solution of buffer + protein.

<sup>h</sup> Average of measured absorbance or calculated concentration values.

<sup>i</sup> Standard deviation of absorbance measurements or calculated concentration.

<sup>j</sup> % Error = (error/average)  $\times$  100%.

From the data collected at 205 nm and 280 nm, and under the assumption that the molar absorptivities calculated for 280 nm are correct, we estimated molar absorptivities at 205 nm for each of the polypeptides in this study, ranging from  $\epsilon_{205} = 96,705 \text{ M}^{-1}\cdot\text{cm}^{-1}$  (37.21 mL·mg<sup>-1</sup>·cm<sup>-1</sup>) for CaMK1<sub>299-320</sub> to  $\epsilon_{205} = 1,918,988 \text{ M}^{-1}\cdot\text{cm}^{-1}$  (30.24 mL·mg<sup>-1</sup>·cm<sup>-1</sup>) for Enzyme I (Table II). Values of  $\epsilon_{205}$  could be calibrated for any protein or peptide containing tryptophan or tyrosine in a similar manner, or using some other method for determining the concentration (e.g., directly weighing lyophilized powder) for those that do not. However, in this study we demonstrate that one can calculate the  $\epsilon_{205}$  directly from the amino acid sequence, making such additional measurements or standards completely unnecessary.

A great advantage of using  $A_{205}$  for determining protein concentration is that absorbance at 205 nm arises primarily from the peptide backbone, although

side chains, and in particular aromatic ones, can contribute significantly. To take this into account, we referenced molar absorptivities at 205 nm given by Goldfarb *et al.*<sup>7</sup> and Saidel *et al.*<sup>12</sup> These reported molar absorptivities for individual amino acids ranged from 54 M<sup>-1</sup>·cm<sup>-1</sup> (glycine) to 20,400 M<sup>-1</sup>·cm<sup>-1</sup>, though only those with  $\epsilon_{205} > 200 \text{ M}^{-1}\cdot\text{cm}^{-1}$  were included in this study (Table III). The literature, however, gives a wide range of values for the  $\epsilon_{205}$  of the peptide bond, and Goldfarb *et al.*<sup>7</sup> suggested a range of 2500–2800 M<sup>-1</sup>·cm<sup>-1</sup> per peptide bond. This likely represents true variation stemming from local structure, rather than experimental error.<sup>7,13</sup>

Due to the above uncertainty, we asked whether we could find an averaged value of  $\epsilon_{205}$  for the peptide bond ( $\epsilon_{bb}$ ) that would be generally applicable, keeping in mind that in reality the actual value per peptide bond is probably quite variable. From our data on eight different polypeptides, we found an overall best-fit  $\epsilon_{bb}$  of  $2780 \pm 168 \text{ M}^{-1}\cdot\text{cm}^{-1}$ , which is

**Table II.** Comparison of Different Methods of Calculating Molar Absorptivity at 205 nm ( $\epsilon_{205}$ )

Protein		$\epsilon_{205}$			
		$A_{280}$ <sup>a</sup>	31 mL·mg <sup>-1</sup> ·cm <sup>-1</sup> <sup>b</sup>	Scopes method <sup>c</sup>	New method <sup>d</sup>
CaMK1 <sub>299-320</sub>	$\epsilon_{205}$ ( $M^{-1}\cdot\text{cm}^{-1}$ ) <sup>e</sup>	96,705	80,572	87,914	98,310
	$\epsilon_{205}$ (mL·mg <sup>-1</sup> ·cm <sup>-1</sup> ) <sup>f</sup>	37.21	31	33.82	37.82
	% Error <sup>g</sup>		-16.68	-9.09	1.66
skMLCK	$\epsilon_{205}$ ( $M^{-1}\cdot\text{cm}^{-1}$ ) <sup>e</sup>	113,397	91,898	97,294	111,950
	$\epsilon_{205}$ (mL·mg <sup>-1</sup> ·cm <sup>-1</sup> ) <sup>f</sup>	38.25	31	32.82	37.76
	% Error <sup>g</sup>		-18.96	-14.20	-1.28
GB1	$\epsilon_{205}$ ( $M^{-1}\cdot\text{cm}^{-1}$ ) <sup>e</sup>	264,682	224,623	228,392	243,070
	$\epsilon_{205}$ (mL·mg <sup>-1</sup> ·cm <sup>-1</sup> ) <sup>f</sup>	36.53	31	31.52	33.55
	% Error <sup>g</sup>		-15.13	-13.71	-8.17
talin2	$\epsilon_{205}$ ( $M^{-1}\cdot\text{cm}^{-1}$ ) <sup>e</sup>	399,873	358,304	370,898	390,810
	$\epsilon_{205}$ (mL·mg <sup>-1</sup> ·cm <sup>-1</sup> ) <sup>f</sup>	34.60	31	32.09	33.81
	% Error <sup>g</sup>		-10.40	-7.25	-2.27
apo CaM	$\epsilon_{205}$ ( $M^{-1}\cdot\text{cm}^{-1}$ ) <sup>e</sup>	514,838	517,889	462,669	524,190
	$\epsilon_{205}$ (mL·mg <sup>-1</sup> ·cm <sup>-1</sup> ) <sup>f</sup>	30.82	31	27.69	31.38
	% Error <sup>g</sup>		0.59	-10.13	1.82
4Ca <sup>2+</sup> -CaM	$\epsilon_{205}$ ( $M^{-1}\cdot\text{cm}^{-1}$ ) <sup>e</sup>	492,873	517,889	463,186	524,190
	$\epsilon_{205}$ (mL·mg <sup>-1</sup> ·cm <sup>-1</sup> ) <sup>f</sup>	29.50	31	27.73	31.38
	% Error <sup>g</sup>		5.08	-6.02	6.35
CaMK1 <sub>1-296</sub>	$\epsilon_{205}$ ( $M^{-1}\cdot\text{cm}^{-1}$ ) <sup>e</sup>	1,111,071	1,036,715	1,057,752	1,155,120
	$\epsilon_{205}$ (mL·mg <sup>-1</sup> ·cm <sup>-1</sup> ) <sup>f</sup>	33.22	31	31.63	34.54
	% Error <sup>g</sup>		-6.69	-4.80	3.96
MBP	$\epsilon_{205}$ ( $M^{-1}\cdot\text{cm}^{-1}$ ) <sup>e</sup>	1,506,589	1,259,043	1,311,224	1,457,280
	$\epsilon_{205}$ (mL·mg <sup>-1</sup> ·cm <sup>-1</sup> ) <sup>f</sup>	37.10	31	32.28	35.88
	% Error <sup>g</sup>		-16.43	-12.97	-3.27
Enzyme I	$\epsilon_{205}$ ( $M^{-1}\cdot\text{cm}^{-1}$ ) <sup>e</sup>	1,918,988	1,967,462	1,810,473	1,973,840
	$\epsilon_{205}$ (mL·mg <sup>-1</sup> ·cm <sup>-1</sup> ) <sup>f</sup>	30.24	31	28.53	31.10
	% Error <sup>g</sup>		2.53	-5.65	2.86

<sup>a</sup> Molar absorptivity at 205 nm calculated directly from  $A_{280}$  data ( $\epsilon_{205} = A_{205} \times \epsilon_{280}/A_{280}$ ).

<sup>b</sup> Molar absorptivity at 205 nm calculated from generic absorptivity of 31 mL·mg<sup>-1</sup>·cm<sup>-1</sup> [ $\epsilon_{205}$  ( $M^{-1}\cdot\text{cm}^{-1}$ ) = 31 (mL·mg<sup>-1</sup>·cm<sup>-1</sup>) × MW (g·mol<sup>-1</sup>)].

<sup>c</sup> Absorptivity at 205 nm first calculated in mass units, as described by Scopes<sup>6</sup> [ $\epsilon_{205}$  (mL·mg<sup>-1</sup>·cm<sup>-1</sup>) = 27.0 + 120 × ( $A_{280}/A_{205}$ )], then transformed into molar units [ $\epsilon_{205}$  ( $M^{-1}\cdot\text{cm}^{-1}$ ) =  $\epsilon_{205}$  (mL·mg<sup>-1</sup>·cm<sup>-1</sup>) × MW (g·mol<sup>-1</sup>)].

<sup>d</sup> Molar absorptivity calculated from the amino acid sequence as described in the main text; an interactive web server to carry out this calculation online is available at <http://spin.niddk.nih.gov/clore>.

<sup>e</sup> Molar absorptivity at 205 nm calculated by various methods.

<sup>f</sup> Absorptivity at 205 nm, given in alternative mass units [ $=\epsilon_{205}$  ( $M^{-1}\cdot\text{cm}^{-1}$ )/MW (g·mol<sup>-1</sup>)].

<sup>g</sup> Percentage of deviation from  $\epsilon_{205}$  value calculated directly from  $A_{280}$  data.

**Table III.** Molar Absorptivity Values at 205 nm ( $\epsilon_{205}$ ) Used in This Study for Protein Side Chains and the Backbone Peptide Bond

Side chain/feature	$\epsilon_{205}$ ( $M^{-1}\cdot\text{cm}^{-1}$ )
Tryptophan	20,400
Phenylalanine	8600
Tyrosine	6080
Histidine	5200
Methionine	1830
Arginine	1350
Cysteine	690
Asparagine <sup>a</sup>	400
Glutamine <sup>a</sup>	400
Cystine <sup>b</sup>	2200
Backbone peptide bond <sup>c</sup>	2780 ± 168

<sup>a</sup> Values for asparagine and glutamine come from Saidel *et al.*<sup>12</sup> All other values are from Goldfarb *et al.*<sup>7</sup>

<sup>b</sup> If the protein has a disulfide bond, add 820  $M^{-1}\cdot\text{cm}^{-1}$  (2200  $M^{-1}\cdot\text{cm}^{-1}$  - 2 × 690  $M^{-1}\cdot\text{cm}^{-1}$ ) to its  $\epsilon_{205}$ .

<sup>c</sup> Best-fit value determined as described in the text and given as the average ± 1 standard deviation.

within the previously reported range.<sup>7</sup> The values calculated for individual peptides ranged from 2567 to 3112  $M^{-1}\cdot\text{cm}^{-1}$  (Table I). Using the average value of  $\epsilon_{\text{bb}} = 2780 M^{-1}\cdot\text{cm}^{-1}$ , we calculated a value of  $\epsilon_{205}$  for each polypeptide in this study by summing all of the backbone and side chain contributions, as described in the methods. These values are presented in the rightmost column of Table I. Compared to the concentrations determined from  $A_{280}$ , the  $\epsilon_{205}$  values based on  $\epsilon_{\text{bb}} = 2780 M^{-1}\cdot\text{cm}^{-1}$  give errors in concentration of at most 8.2% (GB1), though for all but two (GB1 and 4Ca<sup>2+</sup>-CaM) the error is less than 4% (Table II) and well within experimental error (Table I). If, however, one uses a non-sequence-specific estimate for  $\epsilon_{205}$  of 31 mL·mg<sup>-1</sup>·cm<sup>-1</sup>, the errors are in general much larger (up to 19.0%). Even using the formula put forth by Scopes,<sup>6</sup> these errors are only slightly reduced to a maximum of 14.2% (Table II). Only for 4Ca<sup>2+</sup>-CaM do both of the generic methods give lower errors than the new




sequence-specific method, though the differences in error between these methods (6.4% vs. 5.1% vs. 6.0%) are negligible. For apo CaM and Enzyme I, the generic  $31 \text{ mL}\cdot\text{mg}^{-1}\cdot\text{cm}^{-1}$  value gives a slightly lower error than the sequence-specific value, but the difference in errors (1.8% vs. 0.59% for apo CaM, 2.9% vs. 2.5% for Enzyme I) are miniscule and well within the experimental error. In all other cases, the sequence-specific value of  $\epsilon_{205}$  gives the lowest error, often dramatically so.

Based on the data presented here, this method works well for both folded proteins and short unstructured peptides. Although the majority of the  $A_{205}$  in all cases comes from the peptide backbone, the actual contribution for these test cases ranges from 59.4% (CaMK1<sub>299–320</sub>) to 80.8% (Enzyme I). Thus, one can see why calculating sequence-specific molar absorptivities would be advantageous, since side chains can be responsible for up to 40% of the total  $A_{205}$  in these instances. Additionally, measuring data on both apo CaM and  $4\text{Ca}^{2+}$ -CaM offered the opportunity to test whether structural changes within the same protein affect  $A_{205}$ . Although these two conditions gave slightly different results ( $\epsilon_{\text{bb}} = 2716 \pm 324 \text{ M}^{-1}\cdot\text{cm}^{-1}$  for apo CaM and  $2567 \pm 199 \text{ M}^{-1}\cdot\text{cm}^{-1}$  for  $4\text{Ca}^{2+}$ -CaM), these differences were within the experimental error.

To test the utility of the  $A_{205}$  method for determining protein concentration using sequence-specific calculated molar absorptivities, we performed a concentration-dependent experiment on a protein construct that does not absorb at 280 nm. The N-terminal domain of CaM (CaM<sub>1–76</sub>) does not contain any 280 nm-absorbing residues (both tyrosines are located in the C-terminal domain). Using the method described here, CaM<sub>1–76</sub> has a calculated  $\epsilon_{205}$  of  $266,150 \text{ M}^{-1}\cdot\text{cm}^{-1}$ . We determined the concentrations of CaM<sub>1–76</sub> and skMLCK by  $A_{205}$  and then performed a fluorescence anisotropy binding experiment. This titration experiment yielded a CaM<sub>1–76</sub>:skMLCK stoichiometry of  $2.13(\pm 0.17):1$  (Fig. 1), well within the experimental error of the expected value of 2:1.<sup>14</sup> (Note that whereas the stoichiometry of the full-length CaM:skMLCK interaction is 1:1—with the N and C domains of CaM participating in the interaction<sup>11</sup>—in isolation, individual domains of CaM bind with a 2:1 stoichiometry, because a second CaM<sub>1–76</sub> molecule, in this case, takes the place of the absent C-terminal CaM domain in the complex.<sup>14</sup>)

In conclusion, we have demonstrated the utility of a method for determining protein concentrations from absorption at 205 nm, using a molar absorptivity calculated specifically from the amino acid sequence. This method allows one to easily and quickly determine protein concentration when there are no tryptophan or tyrosine residues present. It is generally applicable, requiring no additional



**Figure 1.** Testing the stoichiometry of  $2\text{Ca}^{2+}$ -CaM<sub>1–76</sub> binding to the skMLCK M13 peptide. Fluorescence anisotropy was measured for the tryptophan of skMLCK alone ( $10 \mu\text{M}$ ) and in the presence of  $0$ – $50 \mu\text{M}$  CaM<sub>1–76</sub>. Experimental data (average of three measurements) are plotted as filled-in squares, with error bars indicating one standard deviation. The best-fit line is shown in red. The CaM<sub>1–76</sub>:skMLCK stoichiometry from the fit of the data is  $n = 2.13 \pm 0.17$ , which is in excellent agreement with the literature value of 2:1.<sup>14</sup> The titration was performed with  $[\text{skMLCK}] \gg K_D$  to most accurately determine the stoichiometry. The effective  $K_D$  from the fit is  $100 \pm 28 \text{ nM}$ , several orders of magnitude weaker than the value determined for full-length CaM ( $\sim 50 \text{ pM}$ ).<sup>20</sup>

standards or other calibration, and we have made this tool available on the web (at <http://spin.niddk.nih.gov/clore>) to assist in calculating  $\epsilon_{205}$  for any protein or peptide. We have also shown that taking the amino acid sequence into account is essential for accurate results, and we believe that this method will be widely applicable.

## Methods

### Protein production and sample preparation

All proteins were produced recombinantly in *Escherichia coli* (with the exception of the two peptides that were commercially synthesized as described below). The 148-residue human calmodulin (CaM) protein was expressed and purified as described previously from a construct in a pET21a vector.<sup>9</sup> Experiments were performed on CaM in a buffer consisting of  $25 \text{ mM}$  4-(2-hydroxyethyl)-1-piperazineethanesulfonic acid (HEPES), pH 6.5,  $100 \text{ mM}$  KCl, 0.02% sodium azide,  $1\times$  Roche Complete Protease Inhibitor, and either  $8 \text{ mM}$   $\text{CaCl}_2$ , ( $4\text{Ca}^{2+}$ -CaM) or  $2 \text{ mM}$  ethylenediaminetetraacetic acid and  $2 \text{ mM}$  ethylene glycol tetraacetic acid (apo CaM). A construct corresponding to the N-terminus of CaM (residues 1–76; CaM<sub>1–76</sub>) was generated using the QuikChange II Site-Directed Mutagenesis Kit (Agilent Technologies), and expressed and purified like full-length CaM. The skMLCK M13 peptide (KRRWKKNFIAVSAANRFKIKISSGAL) and the

peptide corresponding to the CaM-binding domain of CaMK1 (residues 299–320; AKSKWKQAFNA-TAVVRHMRKLQ) were commercially synthesized by Anaspec. They were resuspended in water from lyophilized powder for the experiments here.

The B1 domain of streptococcal protein G (66 residues) was expressed from a construct in a GEV2 vector, as previously described.<sup>15</sup> Experiments were performed in a buffer consisting of 50 mM sodium phosphate, pH 6.5, and 100 mM NaCl. The F3 domain of human talin2 (residues 311–408) was expressed from a construct in a pGEX-6P-2 vector, as previously described.<sup>16</sup> Experiments were performed in a buffer consisting of 50 mM sodium phosphate, pH 7.0, and 100 mM NaCl. A mutant of *E. coli* MBP (370 residues; K1A K46C I212C) was expressed from a construct in a pET11 vector, as previously described.<sup>17</sup> Experiments were performed in a buffer consisting of 20 mM Tris, pH 7.4, 100 mM NaCl, and 1 mM dithiothreitol (DTT). A mutant of *E. coli* Enzyme I (residues 1–575; H189A R367K) was expressed from a construct in a pET11 vector, as previously described.<sup>18</sup> Experiments were performed in H<sub>2</sub>O.

DNA corresponding to the full-length 374-residue calcium/calmodulin-dependent protein kinase type 1 (CaMK1) protein from rat was synthesized (codon-optimized for expression in *E. coli*) by Genescript (<http://www.genescript.com>) and subcloned into the pET47b vector. A construct corresponding to residues 1–296 was generated using the QuikChange II kit. CaMK1 was expressed in BL21-Star cells (Agilent Technologies) using standard methods. Briefly, cells were grown at 37°C in 1 L of Luria Bertini (LB) medium to OD<sub>600nm</sub> (optical density) ~ 0.6, cooled to 25°C, and induced with 1 mM isopropyl β-D-1-thiogalactopyranoside (IPTG). Cells were harvested by centrifugation ~16 h later. Cells were resuspended in 50 mM sodium phosphate, pH 7.0, 300 mM NaCl, 1 mM DTT, and 1× Roche Complete Protease Inhibitor. Cells were lysed using a microfluidizer, cleared by centrifugation, and then loaded on a HisTrap column (GE Life Sciences). Protein was eluted with a 0–100% gradient over five column volumes with the same buffer containing 500 mM imidazole. The protein was cleaved ~20 h at 4°C with 3C protease, then further purified by gel filtration on a HiLoad 26/60 Superdex 75 column (GE Life Sciences) into a buffer consisting of 25 mM HEPES, pH 6.5, 100 mM KCl, 0.02% sodium azide, and 0.1× Roche Complete Protease Inhibitor. Fractions were concentrated with Amicon Ultra Centrifugal Filter Units (10 kDa molecular weight cutoff).

3C Protease was expressed as a GST fusion (from a pGEX vector) in BL21 CodonPlus cells (Agilent Technologies). Cells were grown at 37°C in 1 L of LB medium to OD<sub>600nm</sub> ~ 0.6, cooled to 25°C, and induced with 0.2 mM IPTG. Cells were harvested by

centrifugation ~16 h later and resuspended in 50 mM sodium phosphate, pH 7.0, 100 mM NaCl, and 1 mM DTT. Cells were lysed using a microfluidizer, cleared by centrifugation, and then loaded on a GStrap column (GE Life Sciences). Protein was eluted with 30 mM reduced glutathione. Aliquots were stored at –80°C in elution buffer.

### Absorbance measurements

Absorbance values at 205 nm and 280 nm were measured on an Agilent 8453 UV/Vis spectrophotometer, using a quartz cuvette with a 1 cm path length. For all samples, dilutions were performed in water (due to the high UV absorbance of some of the buffers used). Absorbance was measured at various dilutions in half-log increments, and all measurements within a linear range for absorbance versus concentration were averaged (after extrapolating the undiluted absorbance value by multiplying by the dilution factor). The absorbance values of the buffers alone were also measured in the same manner. For all samples, the final  $A_{205}$  used for protein concentration determination was the average extrapolated  $A_{205}$  for the undiluted protein solution minus the  $A_{205}$  for the buffer alone (Table I).

### Calculation of molar absorptivities

Absorbance ( $A_\lambda$ ) at a given wavelength  $\lambda$  is given by the Beer–Lambert law:

$$A_\lambda = \epsilon_\lambda c l \quad (1)$$

where  $\epsilon_\lambda$  is the molar absorptivity at wavelength  $\lambda$ ,  $c$  the concentration, and  $l$  the path length (always 1 cm in this study). For measurements at 280 nm, the  $\epsilon_{280}$  was calculated by adding 5500  $M^{-1}\cdot\text{cm}^{-1}$  for each tryptophan and 1290  $M^{-1}\cdot\text{cm}^{-1}$  for each tyrosine, based on standard literature values.<sup>2–4,19</sup>

For measurements at 205 nm, the absorbance of both the peptide backbone and side chains were taken into account. The values used for each side chain are given in Table III and were taken from literature values (Goldfarb *et al.*<sup>7</sup> for all values, with the exception of the values for glutamine and asparagine, which came from Saidel *et al.*<sup>12</sup>). The molar absorptivity at 205 nm ( $\epsilon_{205}$ ) for a polypeptide is given by the formula:

$$\epsilon_{205} = \sum (\epsilon_i n_i) + \epsilon_{\text{bb}} (r - 1) \quad (2)$$

where for each amino acid type  $i$ ,  $\epsilon_i$  is the molar absorptivity of that amino acid type (Table III),  $n_i$  is the number of times that amino acid type appears in the polypeptide sequence,  $\epsilon_{\text{bb}}$  is the molar absorptivity for a single backbone peptide bond, and  $r$  is the number of residues in the polypeptide sequence.

A standard value for the absorbance of the peptide bond could not be chosen a priori from the literature, so instead one was determined empirically. This was accomplished first by calculating a value of  $\epsilon_{205}$  for each polypeptide in the study:

$$\epsilon_{205} = A_{205} \frac{\epsilon_{280}}{A_{280}} \quad (3)$$

where  $\epsilon_{280}$  is the known molar absorptivity at 280 nm for that polypeptide, and  $A_{205}$  and  $A_{280}$  are the absorbance data for that polypeptide (Table I). These  $A_{280}$ -based  $\epsilon_{205}$  values are presented in the leftmost data column of Table II. For each polypeptide, a value of  $\epsilon_{bb}$  was then calculated by rearrangement of Eq. (2):

$$\epsilon_{bb} = \frac{\epsilon_{205} - \sum (\epsilon_i n_i)}{r - 1} \quad (4)$$

These values of  $\epsilon_{bb}$  are presented in Table I. An overall best-fit value of  $\epsilon_{bb}$  was then computed by averaging the  $\epsilon_{bb}$  values for the individual polypeptides. The use of this single optimized value (determined to be  $2780 \pm 168 \text{ M}^{-1} \cdot \text{cm}^{-1}$  in this case) is not meant to imply that each peptide bond always displays the same absorbance value (since the absorbance of the peptide bond is known to be quite

variable<sup>7,13</sup>) but rather that the use of this one value is sufficiently reliable to be generally useful.

Thus, using a value of  $\epsilon_{bb} = 2780 \text{ M}^{-1} \cdot \text{cm}^{-1}$ ,  $\epsilon_{205}$  can be calculated for any protein or peptide directly from its amino acid sequence, using Eq. (2). For the polypeptides used in this study,  $\epsilon_{205}$  values calculated in this manner are presented in the rightmost column of Table II. A web server that performs this calculation can be accessed online at <http://spin.niddk.nih.gov/clore>. This tool also calculates the molecular weight for various universal isotopic labeling schemes that might be used in nuclear magnetic resonance studies.

### Fluorescence experiments

Fluorescence experiments were carried out at 27°C using a Jobin Yvon FluoroMax-3 fluorometer equipped with a Peltier temperature control unit. The fluorescence anisotropy of the single tryptophan residue in the skMLCK peptide was monitored with excitation at 295 nm and emission at 357 nm. Measurements were acquired on 10  $\mu\text{M}$  skMLCK in the presence of 0–50  $\mu\text{M}$   $2\text{Ca}^{2+}$ -CaM<sub>1-76</sub>. Experiments were performed in 25 mM HEPES, pH 6.5, 100 mM KCl, 8 mM CaCl<sub>2</sub>, and 0.1× Roche Complete Protease Inhibitor. Data were analyzed by fitting the fluorescence anisotropy versus [CaM<sub>1-76</sub>] to the following equation:

$$A = A_{\min} + A_{\max} \left( \frac{n[L] + [U] + K_D^{\text{app}} - \sqrt{(n[L] + [U] + K_D^{\text{app}})^2 - 4n[L][U]}}{2n[L]} \right) \quad (5)$$

where  $A$  is the measured fluorescence anisotropy at each point (the dependent variable in the fitting),  $A_{\max}$  the anisotropy of the skMLCK peptide fully saturated with CaM<sub>1-76</sub>,  $A_{\min}$  the anisotropy of the free skMLCK peptide,  $n$  the stoichiometry of CaM<sub>1-76</sub>:skMLCK binding,  $[L]$  the total (bound + unbound) concentration of the skMLCK peptide,  $[U]$  the total (bound + unbound) concentration of CaM<sub>1-76</sub> at each point (the independent variable in the fitting), and  $K_D^{\text{app}}$  the apparent equilibrium dissociation constant, defined as:

$$K_D^{\text{app}} = \frac{[U]_{\text{unbound}} [L]_{\text{unbound}}}{[U_n L]} \quad (6)$$

$K_D^{\text{app}}$  is only equal to the true dissociation constant for the case where the stoichiometry is  $n = 1$ ; otherwise it represents an effective  $K_D$  (i.e., concentration at half saturation). Experiments were carried out

under conditions where  $[L] \gg K_D$  to most accurately fit the stoichiometry. Data were fit using OriginPro 8.

### Acknowledgments

We thank David Libich for useful discussions and Vincenzo Venditti for providing purified MBP and Enzyme I.

### References

1. Stoscheck CM (1990) Quantitation of protein. *Methods Enzymol* 182:50–68.
2. Edelhoch H (1967) Spectroscopic determination of tryptophan and tyrosine in proteins. *Biochemistry* 6:1948–1954.
3. Gill SC, von Hippel PH (1989) Calculation of protein extinction coefficients from amino acid sequence data. *Anal Biochem* 182:319–326.
4. Pace CN, Vajdos F, Fee L, Grimsley G, Gray T (1995) How to measure and predict the molar absorption coefficient of a protein. *Protein Sci* 4:2411–2423.

5. Grimsley GR, Pace CN (2004) Spectrophotometric determination of protein concentration. *Curr Protoc Protein Sci* Chapter 3:Unit 3 1.
6. Scopes RK (1974) Measurement of protein by spectrophotometry at 205 nm. *Anal Biochem* 59:277–282.
7. Goldfarb AR, Saidel LJ, Mosovich E (1951) The ultraviolet absorption spectra of proteins. *J Biol Chem* 193:397–404.
8. Simonian MH (2004) Spectrophotometric determination of protein concentration. *Curr Protoc Toxicol Appendix 3:A 3G 1–7*.
9. Anthis NJ, Doucleff M, Clore GM (2011) Transient, sparsely populated compact states of apo and calcium-loaded calmodulin probed by paramagnetic relaxation enhancement: interplay of conformational selection and induced fit. *J Am Chem Soc* 133:18966–18974.
10. Goldberg J, Nairn AC, Kuriyan J (1996) Structural basis for the autoinhibition of calcium/calmodulin-dependent protein kinase I. *Cell* 84:875–887.
11. Ikura M, Clore GM, Gronenborn AM, Zhu G, Klee CB, Bax A (1992) Solution structure of a calmodulin-target peptide complex by multidimensional NMR. *Science* 256:632–638.
12. Saidel LJ, Goldfarb AR, Waldman S (1952) The absorption spectra of amino acids in the region two hundred to two hundred and thirty millimicrons. *J Biol Chem* 197:285–291.
13. Rosenheck K, Doty P (1961) The far ultraviolet absorption spectra of polypeptide and protein solutions and their dependence on conformation. *Proc Natl Acad Sci USA* 47:1775–1785.
14. Barth A, Martin SR, Bayley PM (1998) Specificity and symmetry in the interaction of calmodulin domains with the skeletal muscle myosin light chain kinase target sequence. *J Biol Chem* 273:2174–2183.
15. Koenig BW, Rogowski M, Louis JM (2003) A rapid method to attain isotope labeled small soluble peptides for NMR studies. *J Biomol NMR* 26:193–202.
16. Anthis NJ, Wegener KL, Ye F, Kim C, Goult BT, Lowe ED, Vakonakis I, Bate N, Critchley DR, Ginsberg MH, Campbell ID (2009) The structure of an integrin/talin complex reveals the basis of inside-out signal transduction. *EMBO J* 28:3623–3632.
17. Tang C, Schwieters CD, Clore GM (2007) Open-to-closed transition in apo maltose-binding protein observed by paramagnetic NMR. *Nature* 449:1078–1082.
18. Schwieters CD, Suh JY, Grishaev A, Ghirlando R, Takayama Y, Clore GM Solution structure of the 128 kDa enzyme I dimer from *Escherichia coli* and its 146 kDa complex with HPr using residual dipolar couplings and small- and wide-angle X-ray scattering. *J Am Chem Soc* 132:13026–13045.
19. Wilkins MR, Gasteiger E, Bairoch A, Sanchez JC, Williams KL, Appel RD, Hochstrasser DF (1999) Protein identification and analysis tools in the ExPASy server. *Methods Mol Biol* 112:531–552.
20. Grishaev A, Anthis NJ, Clore GM (2012) Contrast-matched small-angle X-ray scattering from a heavy-atom-labeled protein in structure determination: application to a lead-substituted calmodulin-peptide complex. *J Am Chem Soc* 134:14686–14689.